



# National Institute of Medical Device Safety Information

#303-1, Mario Tower, 28, 30-gil, Digital-ro, Guro-gu, Seoul, Korea(08389),  
Tel: +82-2-860-4450, Fax: +82-2-860-4479

## Certificate of Free Sales

No. of Certificate : 20210163962  
Exporting(certifying) country : Republic of Korea  
Importing(requesting) country :

The National Institute of Medical Device Safety Information, certifies that the following firm is authorized to manufacture medical devices under the Medical Device Act, and the following product(s) is(are) permitted to be freely sold in domestic and overseas markets.

○ Applicant (=Product-license holder)

(This certificate shall not be issued to others than the product-license holder)

- Name : iMedisync Inc.

- Address : 5th Fl. 175 Yeoksam-ro, Gangnam-gu Seoul, Korea Rep.

- Registered No : Manufacturer 6263

| No. and date of product-license, comments | Classification   |
|---|--|
| 21-5051 / DEC. 17, 2021                   | Electroencephalograph [2]<br>Name of Product : iSyncWave |

※ Attached, if necessary (approved product information)

☒ Model(Export Name) ☐ Medical Device Accessories ☒ Manufacturer/Legal manufacturer  
☐ Combined/Composite medical device

Issued date : DEC. 31, 2021 (Certificate No.20210163962)

Certified by

Director of Medical Device Certification Department  
National Institute of Medical Device Safety Information

The National Institute of Medical Device Safety Information(NIDS) is related organization of Ministry of Food and Drug Safety and established under Article 42 of the Medical Device Act.  
The Minister of Food and Drug Safety authorized the director of NIDS to issue this free sales certificate under the Article 44 (2) of the Medical device Act.



Attachment

Product License No. : 21-5051 (DEC. 17. 2021)

Classification : Electroencephalograph

Model(Export Name)

· ISW-MKR101



Attachment

Product License No. : 21-5051 (DEC. 17. 2021)

Manufacturer :

Legal manufacturer :

Combined medical device

---

not applicable

Composite medical device

---

not applicable

